# **Science Together**



# Device support in KNAUER CDS / KNAUER device support in 3<sup>rd</sup> party software OpenLAB® CDS EZChrom Edition / ClarityChrom® / Chromeleon™ / ChromGate® / PurityChrom®; other software

The tables list the software package and driver version (in case multiple versions exist); **the version number listed is the** <u>first version that supports the respective device</u>. Newer software versions, of course, also include the driver.

Limitations are marked with superscript, e.g. <sup>1)</sup>. Limitations may affect device functionality, control interfaces or software versions. "n/a" means, that a device is not supported, "-" that no information is available. If a device is supported via different communication ports, for most devices the desired port must be selected in the device directly. From version 7.4.1, ClarityChrom® *Prep* is discontinued; all its prep-functionality is now available in ClarityChrom® with KNAUER FRC license. For some time, ChromGate® was also available as "PLATINblue Edition" for the PLATINblue devices, while the standard version is named as "regular". From driver version 2825 all drivers are available in the regular version; the PLATINblue Edition was discontinued. ChromGate® was discontinued in 2013 at all. OpenLAB® CDS EZChrom Edition version number lists Shared Service version and EZChrom version. KNAUER OEM devices are not supported by the regular KNAUER drivers. At the end of the document other software packages, e.g. KNAUER Mobile Control, are listed. The "Table of contents" includes links for jumping to the desired part of the document. Drivers for OpenLAB® CDS ChemStation Edition or OpenLAB® CDS 2.x are not available.

# Table of contents

| Pumps                       | Z  |
|-----------------------------|----|
| Detectors                   | 4  |
| Assistants                  | 11 |
| Autosamplers                | 12 |
| Analog/Digital Converters   | 15 |
| Thermostats/Column Ovens    | 16 |
| Electrical Valve Drives     | 17 |
| Other Supported Instruments | 18 |
| Fraction Collectors         | 19 |
|                             |    |

| Mobile Control - KNAUER                   | 23 |
|---|----|
| EuroChrom for Windows 3.05                | 24 |
| Other Software                            | 2! |
| OpenLAB CDS (Agilent)                     | 2! |
| OpenLAB CDS ChemStation Edition (Agilent) | 2! |
| LabVIEW (National Instruments)            | 2! |
| HyStar™                                   | 2! |
| Xcalibur™                                 | 2  |

| Pumps  | OpenLAB® CDS<br>EZChrom Edition<br>CDS version and<br>Driver version | ClarityChrom® CDS version and Driver version | Chromeleon™<br>CDS version<br>Driver version                            | ChromGate® CDS version and Driver version               | PurityChrom®      |
|--|--|--|---|---|-------------------|
| AZURA® Pump P 6.1L                                   | OLEE A.01.05 /<br>A.04.05<br>Driver 1022 <sup>14)</sup>              | CC 5.0.2 Driver<br>4095 <sup>1, 15)</sup>    | Cm 7.2 Regular Class<br>3 driver 1024 1)                                | n/a   | Version: 5.07.039 |
| AZURA® Pump P 2.1L                                   | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                             | CC 3.0.7 Driver<br>3078 <sup>1)</sup>        | Cm 7.2 Regular Class<br>3 driver 1024 1)<br>Cm 6.8 Generic<br>driver 2) | CG 3.3.2 Driver 2825                                    | Version: 5.07.039 |
| AZURA® Pump P 2.1S/P<br>4.1S <sup>6)</sup>           | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                             | CC 3.0.7 Driver<br>3078                      | Cm 7.2 Regular Class<br>3 driver 1056 1)<br>Cm 6.8 Generic<br>driver 2) | CG 3.3.2 Driver 1787                                    | Version: 5.07.039 |
| PLATINblue Pump P-1                                  | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                             | CC 5.0.3 Driver<br>4102                      | Cm 6.8 SR10<br>Regular Class 3 driver                                   | CG 3.3.2<br>PLATINblue Edition<br>1751;<br>Regular 2825 | n/a               |
| Smartline Pump 1050 <sup>4)</sup>                    | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                             | CC 2.6.5 Driver<br>1100                      | Cm 6.8 SR10<br>Regular Class 3 driver                                   | CG 3.3.2 Driver 1751                                    | Version: 5.07.039 |
| Smartline Pump 1000 <sup>13)</sup>                   | n/a  | CC 2.4.4                                     | Cm 6.8 Generic driver <sup>2)</sup>                                     | CG 3.1  | Version: 5.07.039 |
| Smartline Pump 100 <sup>5), 6)</sup>                 | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                             | CC 2.6                                       | Cm 6.8 Generic<br>driver <sup>2)</sup>                                  | CG 3.3.1 Driver 1640                                    | Version: 5.07.039 |
| Preparative Pump 1800<br>(also WellChrom K-1800) 13) | n/a  | CC 2.6                                       | n/a   | CG 3.1  | Version: 5.07.039 |
| WellChrom Pump K-1001                                | n/a  | CC 2.6                                       | n/a   | CG 2.55 <sup>7)</sup>                                   | Version: 5.07.039 |

| Pumps (continued)                         | OpenLAB® CDS EZChrom Edition CDS version and Driver version | ClarityChrom® CDS version and Driver version | Chromeleon™<br>CDS version<br>Driver version | ChromGate® CDS version and Driver version | PurityChrom®      |
|---|---|--|--|---|-------------------|
| WellChrom Pump K-1000                     | n/a   | CC 2.6                                       | n/a  | CG 2.55 (not supported in version 3.1) 8) | n/a               |
| WellChrom Pump K-501 9)                   | n/a   | CC 2.6                                       | n/a  | CG 2.8                                    | n/a               |
| WellChrom Pump K-120                      | n/a   | CC 2.6                                       | n/a  | CG 2.8                                    | n/a               |
| BlueShadow Pumps<br>10P/20P <sup>6)</sup> | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                    | CC 3.0.7 Driver<br>3078                      | Cm 6.8 Generic<br>driver <sup>2)</sup>       | CG 3.3.2 Driver 1787                      | Version: 5.07.039 |
| BlueShadow Pump 40P 10)                   | OLEE A.01.05 /<br>A.04.05<br>Driver 1022                    | CC 5.0.2 Driver<br>4095                      | n/a  | n/a                                       | Version: 5.07.039 |
| BlueShadow Pump 80P 10)                   | OLEE A.01.05 /<br>A.04.05<br>Driver 1022                    | CC 5.0.2 Driver<br>4095                      | Cm 6.8 Generic<br>driver <sup>2)</sup>       | n/a                                       | Version: 5.07.039 |
| Kontron Pumps<br>320/322/325              | n/a   | CC 3.0.5 <sup>11)</sup>                      | -  | CG 3.1.7 Driver 446                       | n/a               |
| Kontron Pumps<br>420/422/425              | n/a   | CC 3.0.5 <sup>11)</sup>                      | -  | CG 3.1.7 Driver 446                       | Version: 5.07.039 |
| Kontron Pumps<br>520/522/525              | n/a   | CC 3.0.5 <sup>11)</sup>                      | -  | CG 3.1.7 Driver 446                       | n/a               |

- 1) Constant Pressure Mode / Isobar Mode is <u>not</u> supported
- generic drivers do not support all of the device's features (e.g. no Isobar Mode); only control via RS-232; should also work with Chromeleon™ 7.x
- <sup>3)</sup> support of RS-232 port from driver version 2844
- <sup>4)</sup> for control via RS-232 only baud rate of 115200 is supported
- 5) can be controlled as pump K-120 only via RS-232 without pressure sensor in former software versions
- 6) to set a static IP address a separate tool is required
- from ChromGate® version 3.1 only with firmware version 5.xx or higher; pumps with firmware up to version 4.XX must be configured as K-1000
- <sup>8)</sup> from ChromGate® version 3.1 no pressure information; also K-1001 with firmware up to version 4.XX, no pressure trace

- 9) from firmware version 1.23, model K-500 is not supported
- 10) supported from Clarity version 4.X
- <sup>11)</sup> driver was developed by DataApex
- special RS-232 cable is required
- <sup>13)</sup> for full functionality the KNAUER RS-232 Y-cable is required (KNAUER article number A0755)
- support of switching SSVs during run from driver version 4.2.0.2161
- support of switching SSVs during run from driver version 6021

| <b>Detectors</b> [for all PDA or scanning UV detectors the 3D UV license option is required to acquire and review the scan (3D) data] | OpenLAB® CDS<br>EZChrom Edition<br>CDS version and<br>Driver version | ClarityChrom® CDS version and Driver version                        | Chromeleon™<br>CDS version<br>Driver version                                    | ChromGate®<br>CDS version and<br>Driver version | PurityChrom®      |
|---|--|---|---|---|-------------------|
| AZURA® Detector UVD 2.1L  | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                             | CC 3.0.7 Driver<br>3078   | Cm 7.2 Regular Class<br>3 driver 1024<br>Cm 6.8 Generic<br>driver <sup>5)</sup> | CG 3.3.2 Driver 2825                            | Version: 5.08.034 |
| AZURA® Detector MWD<br>2.1L   | OLEE A.01.05 /<br>A.04.05<br>driver 2068                             | CC 5.0.5 Driver 5011  | Cm 7.2 Regular Class<br>3 driver 1024   | n/a   | Version: 5.07.067 |
| AZURA® Detector UVD 2.1S  | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                             | CC 3.0.7 Driver<br>3055   | Cm 7.2 Regular Class<br>3 driver 1056   | CG 3.3.2 Driver 1787                            | Version: 5.07.039 |
| AZURA® Detector DAD 6.1L  | OLEE A.01.05 /<br>A.04.05<br>Driver 2047                             | CC 5.0.5 Driver<br>4110   | Cm 7.2 Regular Class<br>3 driver 1024   | n/a   | Version: 5.07.102 |
| AZURA® Detector DAD 2.1L  | OLEE A.01.05 /<br>A.04.05<br>Driver 2068                             | CC 5.0.5 Driver<br>5011   | Cm 7.2 Regular Class<br>3 driver 1024   | n/a   | Version: 5.07.102 |
| AZURA® Detector RID 2.1L /<br>RID 2.1L Highflow   | OLEE A.01.05 /<br>A.04.05<br>Driver 2092                             | CC 6.0.1 Driver<br>5037 (compatible<br>with CC/CCP 5.x<br>versions) | Cm 7.2 Regular Class<br>3 driver 1038   | n/a   | Version: 5.08.024 |

| Detectors (continued)                                | OpenLAB® CDS EZChrom Edition CDS version and Driver version | ClarityChrom® CDS version and Driver version            | Chromeleon™<br>CDS version<br>Driver version | ChromGate® CDS version and Driver version                | PurityChrom®      |
|--|---|---|--|--|-------------------|
| AZURA® Conductivity<br>Monitor CM 2.1S <sup>2)</sup> | OLEE A.01.05 /<br>A.04.05<br>Driver 1022                    | CC 5.0.2 Driver<br>4095                                 | n/a  | Generic Driver June 2013 <sup>5)</sup>                   | Version: 5.07.039 |
| AZURA® ECD 2.1                                       | n/a   | Supported as Antec<br>Decade Elite;<br>CC 5.X, USB, LAN | Cm 7.2 - coming soon                         | n/a  | n/a               |
| PLATINblue Detector MW-1                             | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                    | CC6.1.0; Driver<br>5015                                 | Cm 6.8 SR10<br>Regular Class 3 driver        | CG 3.3.2<br>PLATINblue Edition<br>1751;<br>Regular 2825  | n/a               |
| PLATINblue DAD PDA-1 <sup>23)</sup>                  | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                    | CC 5.0.2 Driver<br>4095 <sup>4)</sup>                   | Cm 6.8 SR10<br>Regular Class 3 driver        | CG 3.3.2;<br>PLATINblue Edition<br>1751;<br>Regular 2825 | n/a               |
| Smartline UV Detector 2500                           | n/a   | CC 2.4.4  | Cm 6.8 Generic<br>driver <sup>5)</sup>       | CG 3.1   | Version: 5.07.039 |
| Smartline UV Detector 2520                           | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                    | CC 3.0.5 Driver<br>3033                                 | n/a  | CG 3.3.2 Driver 1751                                     | n/a               |
| Smartline UV Detector 2550 3), 7), 23)               | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                    | CC 2.6.5 Driver<br>1079                                 | Cm 6.8 SR10<br>Regular Class 3 driver        | CG 3.3.1 Driver 1640                                     | n/a               |
| Smartline UV Detector 2600                           | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                    | CC 2.4.4  | Cm 6.8 Generic<br>driver <sup>5)</sup>       | CG 3.1   | Version: 5.07.039 |
| Smartline UV Detector 200                            | n/a   | CC 2.6  | n/a  | CG 3.1.7 Driver 1466                                     | n/a               |
| Smartline PDA Detector 2800 9), 23)                  | n/a   | CC 2.4.4  | n/a  | CG 3.1.7 Driver 446                                      | Version: 5.07.039 |
| Smartline PDA Detector<br>2850 <sup>23)</sup>        | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                    | CC 2.6.5 Driver<br>1100                                 | n/a  | CG 3.3.2 Driver 1751                                     | n/a               |

| Detectors (continued)                               | OpenLAB® CDS EZChrom Edition CDS version and Driver Version | ClarityChrom® CDS version and Driver Version                            | Chromeleon™<br>CDS version<br>Driver version | ChromGate® CDS version and Driver Version | PurityChrom® |
|---|---|---|--|---|--------------|
| Smartline RI Detector 2300                          | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                    | CC 2.4.4  | n/a  | CG 3.1                                    | n/a          |
| Smartline RI Detector 2400                          | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                    | CC 2.4.4  | n/a  | CG 3.1                                    | n/a          |
| Smartline Conductivity<br>Monitor S2900             | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                    | CC 3.0.7 Driver<br>3055   | n/a  | Generic Driver<br>September 2010          | n/a          |
| WellChrom UV Detector<br>K-2000 <sup>8)</sup>       | n/a   | CC 2.6  | n/a  | CG 2.8                                    | n/a          |
| WellChrom UV Detector<br>K-2001                     | n/a   | CC 2.6  | n/a  | CG 2.8                                    | n/a          |
| WellChrom UV Detector<br>K-2500 <sup>8)</sup>       | n/a   | CC 2.6  | n/a  | CG 2.8                                    | n/a          |
| WellChrom UV Detector<br>K-2501 <sup>6)</sup>       | n/a   | CC 2.6  | n/a  | CG 2.8                                    |              |
| WellChrom UV Detector<br>K-2600 <sup>22), 23)</sup> | OLEE A.01.05 /<br>A.04.05<br>Driver 2026                    | CC 2.6  | Cm 6.8 Generic<br>driver <sup>5)</sup>       | CG 2.55                                   | n/a          |
| WellChrom PDA Detector<br>K-2700 <sup>9), 23)</sup> | n/a   | PCI with RS-422 (PCI card driver supports up to Windows 7 32-bit) / LAN | n/a  | CG 2.55, LAN from 3.1                     | n/a          |
| WellChrom PDA Detector<br>K-2800 <sup>9), 23)</sup> | n/a   | PCI with RS-422 (PCI card driver supports up to Windows 7 32-bit) / LAN | n/a  | CG 2.8, LAN from 3.1                      | n/a          |
| WellChrom RI Detector<br>2300                       | n/a   | CC 2.6  | n/a  | CG 2.8                                    | n/a          |

| Detectors (continued)   | OpenLAB® CDS<br>EZChrom Edition<br>CDS version and<br>Driver Version  | ClarityChrom® CDS version and Driver Version                               | Chromeleon™<br>CDS version<br>Driver version   | ChromGate® CDS version and Driver Version | PurityChrom®      |
|---|---|--|--|---|-------------------|
| WellChrom RI Detector<br>2301   | n/a   | CC 2.6   | n/a  | CG 2.8                                    | n/a               |
| WellChrom RI Detector<br>2400   | n/a   | CC 2.6   | n/a  | CG 2.8                                    | n/a               |
| WellChrom RI Detector<br>2401   | n/a   | CC 2.6   | n/a  | CG 2.8                                    | n/a               |
| Shimadzu fluorescence<br>detector RF-10Axl; onboard<br>COM or PCI/PCIe COM<br>board required <sup>10)</sup>   | n/a   | CC 2.6   | n/a  | CG 3.1                                    | n/a               |
| Shimadzu fluorescence<br>detector RF-20A/Axs serial<br>control with special<br><b>firmware version 0.90</b> only<br>from KNAUER ( <u>limited</u><br><u>functionality</u> ); onboard COM<br>or PCI/PCIe COM board<br>required <sup>11)</sup> | OLEE A.01.05 /<br>A.04.05<br>Driver 1017  | CC 3.0.7 Driver<br>3055  | n/a  | CG 3.3.2;<br>Driver 1787                  | Version: 5.08.024 |
| Shimadzu fluorescence<br>detector RF-20A/Axs with<br>CBM-20A only (CBM-40<br>does not work with the<br>detector only; Shimadzu<br>pump required)  | Shimadzu Control<br>with official<br>Shimadzu driver;<br>requires special<br>license, driver not<br>developed by<br>KNAUER <sup>19)</sup> | CC 7.2.0, only LC-<br>10 functionality,<br>driver developed by<br>DataApex | Cm 7.2 with official<br>Shimadzu driver;<br>requires special<br>license, driver not<br>developed by<br>KNAUER <sup>20)</sup> | n/a                                       | n/a               |
| Virtual Detector  | OLEE A.01.05 /<br>A.04.05<br>Driver 1017  | CC 2.6.5 Driver<br>1079  | n/a  | CG 3.1.6                                  | n/a               |
| BlueShadow Detector 10D 3)  | OLEE A.01.05 /<br>A.04.05<br>Driver 1017  | CC 3.0.7 Driver<br>3055  | n/a  | CG 3.3.2<br>Driver 1787                   | Version: 5.07.039 |

| Detectors (continued)                                     | OpenLAB® CDS<br>EZChrom Edition<br>CDS version and<br>Driver Version | ClarityChrom® CDS version and Driver Version | Chromeleon™<br>CDS version<br>Driver version   | ChromGate® CDS version and Driver Version                                 | PurityChrom®      |
|---|--|--|--|---|-------------------|
| BlueShadow Detector 40D <sup>3),</sup>                    | OLEE A.01.05 /<br>A.04.05<br>driver 1022                             | CC 5.0.2 Driver<br>4095                      | n/a  | n/a   | n/a               |
| BlueShadow Detector 50D <sup>3),</sup> <sup>7), 23)</sup> | OLEE A.01.05 /<br>A.04.05<br>driver 1022                             | CC 5.0.2 Driver<br>4095                      | n/a  | n/a   | Version: 5.09.134 |
| Alltech Conductivity Detector 650                         | n/a  | n/a  | n/a  | CG 3.1.7 Driver 446   | n/a               |
| Kontron UV Detector<br>332/335 <sup>8), 12)</sup>         | n/a  | supported by DataApex; CC 5.0.2              | -  | CG 3.1.7 Driver 446   | n/a               |
| Kontron UV Detector<br>430/430A/432/435 8), 12)           | n/a  | supported by DataApex; CC 5.0.2              | -  | CG 3.1.7 Driver 446   | n/a               |
| Kontron UV Detector 530/532/535 8), 12)                   | n/a  | supported by DataApex; CC 5.0.2              | -  | CG 3.1.7 Driver 446   | n/a               |
| Kontron DAD 540(/545) 12), 23)                            | n/a  | supported by DataApex; CC 5.0.2              | -  | CG 3.1.7 Driver 1466  | n/a               |
| Sedere Sedex ELSD 80/85LT                                 | supported, driver<br>was developed by<br>MayLab Analytical           | supported by<br>DataApex; CC 3.0.5           | Cm 7.2: 85LT<br>supported from FW<br>2.7, requires Class 3<br>license (driver not<br>developed by<br>KNAUER) | supported, driver was<br>developed by<br>MayLab Analytical <sup>14)</sup> | n/a               |
| Sedere Sedex ELSD 90LT                                    | supported, driver<br>was developed by<br>MayLab Analytical           | supported by<br>DataApex; CC 5.0.2           | Cm 7.2: from FW 1.0,<br>requires Class 3<br>license (driver not<br>developed by<br>KNAUER)                   | supported, driver was<br>developed by<br>MayLab Analytical <sup>14)</sup> | n/a               |
| Sedere Sedex ELSD 100LT                                   | supported, driver<br>was developed by<br>MayLab Analytical           | supported by<br>DataApex; CC 5.0.2           | Cm 7.2: requires<br>Class 3 license (driver<br>not developed by<br>KNAUER)                                   | n/a   | n/a               |

| Detectors (continued)                 | OpenLAB® CDS EZChrom Edition CDS version and Driver Version | ClarityChrom® CDS version and Driver Version  | Chromeleon™<br>CDS version<br>Driver version   | ChromGate® CDS version and Driver Version   | PurityChrom®      |
|---------------------------------------|---|---|--|---|-------------------|
| Sedere Sedex LC                       | supported, driver<br>was developed by<br>MayLab Analytical  | supported by<br>DataApex; CC 5.0.2            | Cm 7.2: From FW<br>A.1.5, requires Class<br>3 license (driver not<br>developed by<br>KNAUER) | n/a   | Version: 5.07.039 |
| ERC RefractoMax 520 series            | n/a   | supported by<br>DataApex; CC 6.1.0,<br>RS-232 | Cm 7.2 (driver not developed by KNAUER)  | n/a   | n/a               |
| Antec Leyden LC-EC<br>Detector Decade | n/a   | supported by<br>DataApex; CC 5.0.2            | Cm 6.8 (driver not developed by KNAUER)  | n/a   | Version: 5.07.039 |
| Antec Decade II                       | n/a   | supported by<br>DataApex; CC 5.x,<br>RS-232   | n/a  | n/a   | n/a               |
| Antec Decade Elite                    | n/a   | supported by<br>DataApex;CC 5.X,<br>USB, LAN  | n/a  | n/a   | n/a               |
| Shimadzu CDD-10Avp <sup>26)</sup>     | n/a   | n/a   | n/a  | n/a   | n/a               |
| Grace/Alltech ELSD 3300               | -   | supported by<br>DataApex; CC 5.x,<br>LAN      | -  | Separate, license-free<br>driver upon request<br>or from Grace<br>webpage, driver not<br>developed by<br>KNAUER | -                 |

| Detectors (continued)  | OpenLAB® CDS EZChrom Edition CDS version and Driver Version | ClarityChrom® CDS version and Driver Version                          | Chromeleon™<br>CDS version<br>Driver version                                    | ChromGate® CDS version and Driver Version | PurityChrom® |
|--|---|---|---|---|--------------|
| Berthold HERM LB500 (also with fLumo)  | n/a   | Ruby script<br>available upon<br>request, USB                         | Cm 7.2: requires<br>Class 3 license (driver<br>not developed by<br>KNAUER), USB | n/a                                       | n/a          |
| Elysia-raytest radioactivity<br>detectors and monitors<br>Gabi Nova, Gabi *, POMO,<br>Ramona * | -   | only upon request<br>to Elysia; no driver<br>from KNAUER<br>available | -   | n/a                                       | n/a          |

- 1) only LAN is supported
- <sup>2)</sup> to set a static IP address a separate tool is required
- <sup>3)</sup> via RS-232 the maximum data rate is limited to 10 Hz
- 4) only in ClarityChrom® 5.0.2: Runs with data rates >10 Hz may not be stored with 3D data; recommended maximum data rate is 10 Hz, fixed in later versions from 5.0.3
- generic drivers do not support all of the device's features; only control via RS-232, <u>data acquisition requires A/D converter</u>; should also work with Chromeleon™ 7.x
- 6) a trigger on Start input also starts the internal wavelength program
- <sup>7)</sup> for control via RS-232 only baud rate 115200 is supported
- <sup>8)</sup> device only supports data acquisition via RS-232, no wavelength control
- <sup>9)</sup> LAN only via separate LAN adapter A4442; default communication adapter is PCI interface card for RS-422, PCI interface card requires a PCI slot on the computer; driver for PCI interface card only works with Windows NT/2000, Windows XP or Windows 7 32-bit. Older ISA interface card is not supported anymore.
- RS-232 only, onboard COM port or PCI/PCIe extension cards are recommended, USBtoCOM adapter and other adapters may not work; maximum data rate 0.5 Hz, A/D converter for higher data rates required; special RS-232 cable required (KNAUER part number G0743)
- RS-232 only, onboard COM port or PCI/PCIe extension cards are recommended, USBtoCOM adapter and other adapters may not work; special firmware version 0.90 required, feature set of RF-10 AXL only (1 channel only, no flow cell temperature control, no scan), maximum data rate 5 Hz via RS-232, A/D converter for higher data rates required; special RS-232 cable required (KNAUER part number G0743)
- 12) special RS-232 cable is required
- data rate limited to 10 Hz, driver must be installed separately and is protected with a separate USB dongle, detector cannot be used for fractionation with KNAUER FRC option; KNAUER part number A2618-01

- data rate limited to 10 Hz, driver must be installed separately and is protected with a separate USB dongle, detector cannot be used for fractionation with KNAUER FRC option; KNAUER part number A1513-07
- <sup>15)</sup> 4 channels, up to 50 Hz data rate
- 8 channels, up to 10 Hz; PurityChrom® offers only limited 3D functionality, refer to the PurityChrom® manual
- up to 2 Hz, not in a system with CM 2.1S (only RID 2.1L or CM 2.1S in a system with PurityChrom® is allowed)
- <sup>18)</sup> not in a system with RID 2.1L (only CM 2.1S or RID 2.1L in a system with PurityChrom® is allowed)
- <sup>19)</sup> requires separate license, KNAUER part number A2618-02, detector cannot be used for fractionation with KNAUER FRC option
- <sup>20)</sup> requires separate license, KNAUER part number A1783-3
- <sup>21)</sup> requires separate license, KNAUER part number A1783-4
- <sup>22)</sup> for full functionality the KNAUER RS-232 Y-cable is required (KNAUER article number A0755)
- <sup>23)</sup> for acquiring 3D data (spectra) a 3D UV (PDA) license option is required.
- not in a system with IFU 2.1/IFU 2.1LAN (only RF-20/Axs or IFU 2.1/IFU 2.1LAN in a system with PurityChrom® is allowed)
- <sup>25)</sup> extended dynamic range for RID 2.1L is not supported in PurityChrom
- <sup>26)</sup> the Shimadzu CDD-10Avp requires an A/D converter for data acquisition, device control is not possible
- <sup>27)</sup> in PurityChrom it is not possible to run UVD 2.1S, UVD 2.1L, MWD 2.1L and 50D in any combination; only one of these device types are allowed in a system (also more than one device of the same type)

| Assistants (up to three devices in one housing) 1) Included valves as fraction collector - see Fraction Collectors | OpenLAB® CDS<br>EZChrom Edition<br>CDS version and<br>Driver Version | ClarityChrom®<br>CDS version and<br>Driver Version | Chromeleon™<br>CDS version<br>Driver version        | ChromGate®<br>CDS version and<br>Driver Version | PurityChrom®      |
|--|--|--|---|---|-------------------|
| ASM 2.1L 1), 2)  | OLEE A.01.05 /<br>A.04.05<br>Driver 1017 <sup>3)</sup>               | CC 3.0.7; Driver 3078 <sup>4) 5) 6)</sup>          | Cm 7.2 Regular Class<br>3 driver 1056 <sup>8)</sup> | CG 3.3.2; Driver 2825                           | Version: 5.07.039 |
| ASM 2.2L 1), 2)  | OLEE A.04.06<br>Driver 2184 <sup>10)</sup>                           | CC 8.2.2; Driver 6036 4) 6)                        | Coming soon   | n/a   | Version: 5.09.036 |

<sup>-</sup> configuration depends on order; refer the allowed configuration and restrictions in our FAQ (https://www.knauer.net/en/Support/FAQ/Assistants)

<sup>&</sup>lt;sup>2)</sup> - only one LAN port for all devices

- only one valve as fraction collector, no cascading if controlled as ASM; all included modules also can be controlled as separate devices via IP port; FRC valve cascading only is possible if a valve in the ASM is configured as KNAUER MultiValve FC and is controlled as separate device via IP port; more information about control as separate device via IP port in software manual
- <sup>4)</sup> ASM is only supported as complete Assistant; included modules <u>cannot</u> be controlled as separate devices (IP port separation is not supported)
- two pumps can be controlled either as HPG or as separate pumps from hotfix 5043, in older versions only as HPG
- 6) only one valve as fraction collector, no cascading
- <sup>7)</sup> ASM is not supported as device; the included modules can be controlled as separate devices via IP port; only one valve as fraction collector
- 8) ASM is not supported as device; the included modules can be controlled as separate devices via IP port; no valve as FRC
- 9) ASM's Event Outs and Start In are not supported
- <sup>10)</sup> Modules of ASM 2.2L must be configured as separate devices

| Autosamplers (one autosampler in a system)       | OpenLAB® CDS EZChrom Edition CDS version and Driver Version | ClarityChrom® CDS version and Driver Version                 | Chromeleon™<br>CDS version<br>Driver version  | ChromGate®<br>CDS version and<br>Driver Version                          | PurityChrom®      |
|--|---|--|---|--|-------------------|
| AZURA® Autosampler<br>AS 6.1L <sup>1), 2)</sup>  | OLEE A.01.05 /<br>A.04.05<br>Driver 2109                    | CC 7.2.0.5043 <sup>6)</sup>                                  | Cm 7.2 Regular Class<br>3 driver 1038 <sup>6)</sup><br>(requires FW 1.17 or<br>higher)                            | n/a  | Version: 5.08.024 |
| Smartline Autosampler 3950 <sup>1), 2)</sup>     | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                    | CC 2.4.4 / LAN from<br>CC 2.6.5 Driver<br>1083 <sup>6)</sup> | Cm 7.2 Regular Class<br>3 driver 1024<br>(requires FW 1.15 or<br>higher)<br>Cm 6.8 SR10<br>Regular Class 3 driver | CG 3.1.7 Driver 1466<br>/ LAN from CG 3.3.1<br>Driver 1640 <sup>8)</sup> | Version: 5.07.067 |
| PLATINblue Autosampler<br>AS-1 <sup>1), 2)</sup> | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                    | CC 6.1.0, Driver<br>5015                                     | Cm 6.8 SR10<br>Regular Class 3 driver   | CG 3.3.2<br>PLATINblue Edition<br>1751 8);<br>Regular 2825 8)            | n/a               |

| Autosampler (continued)  | OpenLAB® CDS EZChrom Edition CDS version and Driver Version | ClarityChrom® CDS version and Driver Version | Chromeleon™<br>CDS version<br>Driver version | ChromGate® CDS version and Driver Version | PurityChrom®      |
|--|---|--|--|---|-------------------|
| Autosampler KNAUER<br>Optimas  | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                    | CC 3.0.7 Driver 3055 <sup>3)</sup>           | n/a <sup>9)</sup>                            | n/a <sup>7)</sup>                         | Version: 5.08.034 |
| Smartline Autosampler<br>3800 (also WellChrom K-<br>3800, Spark Basic<br>Marathon) | n/a   | CC 2.4.4 <sup>3)</sup>                       | Cm 6.8;<br>Generic driver <sup>4)</sup>      | CG 3.3.2 Driver 1787                      | Version: 5.07.039 |
| Smartline Autosampler<br>3900 (also Spark Midas)                                   | n/a   | CC2.4.4 3)                                   | Cm 6.8;<br>Generic driver 4)                 | CG 2.8                                    | Version: 5.07.039 |
| Spark Triathlon  | n/a   | CC 2.4.4 3)                                  | Cm 6.8;<br>Generic driver <sup>4)</sup>      | CG 2.55                                   | Version: 5.07.039 |
| Kontron Autosampler 360  | n/a   | CC 5.0.2 3)                                  | -  | CG 3.1.7 Driver 1466                      | n/a               |
| Kontron Autosampler 46x  | n/a   | -  | -  | CG 3.1.7 Driver 1466                      | n/a               |
| Kontron Autosampler 56x  | n/a   | CC 5.0.2 3)                                  | -  | CG 3.1.7 Driver 1466                      | n/a               |

<sup>-</sup> some device features have been added from time to time; therefore, not all of the current features are supported in the software from beginning as well as not all autosampler firmware versions support it:

#### **AS 6.1L**

- FW 1.22 additional control options for µl-pickup (currently not supported by drivers)
- FW 1.21 1000 μl syringe supported (currently only supported in ClarityChrom), User Defined Program supported (currently not supported by drivers)
- FW 1.20 no driver-dependent changes (extra needle dry step, other needle height during first transport aspiration)
- FW 1.19 100 μl syringe supported: OLEE: driver 2165, CC: 8.1 driver 6034, CG: AS 6.1L not supported; Cm: n/a
- FW 1.18 30 vial tray, 2500  $\mu$ l syringe can be selected also if prep mode is not enabled: OLEE: driver 2143, CC: 7.4.2 driver 5066, CG: AS 6.1L not supported; Cm: 7.x driver 1056
- Advanced needle wash requires FW 1.18: OLEE: driver 2158; CC: 8.1 driver 6018; CG: AS 6.1L not supported; Cm: not supported

#### AS3950

- FW 1.18 2500 μl syringe selectable also if prep mode is not active, 30 vials tray: OLEE: driver 4.2.0.2143; CC: 7.4, driver 5066; Cm: 7.x driver 1056; CG: n/a; PC: 05.09.028 2500 μl syringe; 5.09.069 30 vial tray; 30 vial tray selectable in prep mode: OLEE: 4.2.0.2159; CC: 8.0.1.6012; Cm: 7.x driver 1056; CG: n/a; PC: 5.09.069
- FW 1.16 108 vial tray (Q1 2017); drivers set also 250 μl syringe and 108 vial tray as default in new method: OLEE: driver 2109; CC: 7.2 driver 5045; Cm: 7.x driver 1038; CG: n/a; PC: 05.08.034
- FW 1.15 0.1 μl increments, tray thermostatted up to 40 °C (2014): OLEE: A.04.05 driver 1028; CC: 5.0.3 driver 4102, Cm: 7.x, driver 1024; CG: 3.3.2 driver 2846 (40 °C only); PC: 5.07.067
- FW 1.04 84+3 vial tray (Aug 2011): OLEE: driver 1017; CC: 3.0.7 driver 3055; Cm: 7.x driver 1024; CG: 3.3.2 driver 1787; PC: 5.07.067
- TCP/IP support (is required to control more than one autosampler in a network, devices sold from Q3 2016 with TCP by default, former devices can be switched to TCP): OLEE: driver 1017; CC: 6.1.0 driver 5028 (hotfix); Cm: 7.x driver 1024; CG: n/a (not supported); PC: 5.07.067 (UDP not supported)
- LAN connection (devices with LAN connector sold from 2009): OLEE: driver 1017; CC: 2.6.5 driver 1094; Cm: 7.x driver 1024; CG: 3.3.1 driver 1640; PC: 5.07.067

#### AS-1

- FW 1.03 500 μl syringe is supported
- FW 1.02 Tray thermostatted up to 40 °C: OLEE: from first version; CC: 6.1.0 driver 5015; Cm: 6.8 only; CG: 3.3.2 driver 2846; PC: AS-1 not supported
- 2) LAN default configuration is LAN via DHCP, static IP address requires access to built-in LAN module via browser or ASM software
- <sup>3)</sup> driver was developed by DataApex
- 4) upon request driver was not developed by KNAUER
- 5) special RS-232 cable is required
- <sup>6)</sup> no Stacked Injection supported
- 7) may run with Midas/AS3900 driver
- <sup>8)</sup> in ChromGate TCP/IP communication is not supported, device must be set to UDP; with UDP only one autosampler via LAN in the network can be controlled; later added features (see <sup>1)</sup>) are not supported
- 9) generic driver for autosampler Spark Midas may work; upon request driver was not developed by KNAUER

| Analog/Digital Converters                         | OpenLAB® CDS<br>EZChrom Edition<br>CDS version and<br>Driver Version 1) | ClarityChrom® CDS version and Driver Version                            | Chromeleon™<br>CDS version<br>Driver version | ChromGate® CDS version and Driver Version                          | PurityChrom®      |
|---|---|---|--|--|-------------------|
| KNAUER Interface Box IFU<br>2.1 LAN <sup>2)</sup> | OLEE A.01.05 /<br>A.04.05<br>Driver 2092                                | CC 6.0.1 Driver<br>5043 (compatible<br>with CC/CCP 5.x<br>versions) 15) | n/a  | n/a  | Version: 5.08.024 |
| KNAUER Interface Box IFU 2.1 3), 4), 5)           | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                                | CC 3.0.7 Driver<br>3078, in former<br>version as IF2 5), 15)            | n/a  | CG 3.3.2 Driver 2825,<br>in former version as<br>IF2 <sup>5)</sup> | Version: 5.07.039 |
| KNAUER Interface Box IF2 4)                       | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                                | CC 2.4.4 / 2.6 <sup>7), 8), 15)</sup>                                   | n/a  | CG 3.1 <sup>9)</sup>   | Version: 5.07.039 |
| Smartline Manager 5000 4)                         | n/a   | CC 2.4.4/2.6 7), 8), 15)  | n/a  | CG 3.1 <sup>9)</sup>   | n/a               |
| Smartline Manager 5050 <sup>4)</sup>              | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                                | CC 2.6.5 Driver 1100 <sup>5), 15)</sup>                                 | n/a  | CG 3.3.2 Driver 1751   | n/a               |
| KNAUER HPLC Box Model<br>96 <sup>10)</sup>        | n/a   | CC 2.6 <sup>11)</sup>   | n/a  | CG 2.55 <sup>12), 13)</sup>  | n/a               |
| Kontron Interface Card 14)                        | n/a   | -   | -  | CG 3.1.7 Driver 1466   | n/a               |

<sup>1) -</sup> A/D converter in OpenLAB® must be configured separately; menu is available via Start - Programs - Agilent Technologies - OpenLAB® CDS EZChrom Edition, right click on Agilent SS420X Interface Configuration and select Run as Administrator

www.knauer.net

<sup>2) -</sup> to set a static IP address a 3<sup>rd</sup> party tool is required

<sup>-</sup> built-in USB-to-serial converter requires separate installation of a driver in Windows operating system, can be found on the HPLC software CD or KNAUER webpage (*Download* section for IFU 2.1, Article Number AZB00); check Windows Device Manager for the COM port number of the converter, this number must be selected in the software configuration

<sup>-</sup> as the required serial number in the configuration, enter the last 5 digits of the serial number on the label; if the first digit is a 0 (zero), only enter the last four digits

<sup>&</sup>lt;sup>5)</sup> - in previous software versions IFU 2.1 can be controlled as IF2 or Manager 5000 if IFU 2.1 is not listed in the module list

<sup>6) -</sup> analog input only

<sup>-</sup> in ClarityChrom® version 2.4.4 only data acquisition via analog input is supported (max. 10 Hz), analog out supported from ClarityChrom® version 2.6

- <sup>8)</sup> 50 Hz option for channel 1 supported from version ClarityChrom® version 2.6
- 9) 50 Hz option for channel 1 supported from version ChromGate® version 3.1.4
- as the required serial number in the configuration, enter the last 5 digits of the serial number on the label, if the serial number is higher than 40000, please enter only the last four digits
- supported as "Basic Edition", ClarityChrom® will switch the box into the required mode
- supported as "Integration Box" only for data acquisition via analog input; box must be switched into the correct mode with special software
- <sup>13)</sup> 100 Hz option for channel 1 supported from version ChromGate® version 3.1
- requires an ISA slot on the computer; is not supported from Windows Vista (and higher)
- digital output (one per channel) cannot be programmed; is used automatically for AZ (autozero) signal

| Thermostats/Column Ovens   | OpenLAB® CDS<br>EZChrom Edition<br>CDS version and<br>Driver Version | ClarityChrom®<br>CDS version and<br>Driver Version | Chromeleon™<br>CDS version<br>Driver version | ChromGate®<br>CDS version and<br>Driver Version         | PurityChrom®            |
|--|--|--|--|---|-------------------------|
| Column Thermostat CT 2.1   | OLEE A.01.05 /<br>A.04.05<br>Driver 1027                             | CC 5.0.3 driver<br>4102                            | Cm 7.2 Regular Class<br>3 driver 1024        | n/a   | Version 5.09.069 4), 5) |
| Support of modified CT 2.1 with S/N FCC  | OLEE A.04.06<br>Driver 2184  | CC 8.2.3 driver 6042                               | Cm 7.2 Regular Class<br>3 driver 1056        | n/a   | Version 5.09.134        |
| PLATINblue Column<br>Thermostat T-1  | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                             | CC 5.0.2<br>Driver 4095                            | Cm 6.8 SR10<br>Regular Class 3 driver        | CG 3.3.2<br>PLATINblue Edition<br>1751;<br>Regular 2825 | n/a                     |
| Smartline Column Oven <sup>2)</sup> / Column Thermostat JetStream <sup>1), 2)</sup> / PLATINblue T-1 Basic <sup>2)</sup> | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                             | CC 2.4.4   | Cm 6.8 SR10<br>Regular Class 3 driver        | CG 3.1  | Version 5.07.039        |

| Thermostats/Column Ovens (continued) | OpenLAB® CDS<br>EZChrom Edition<br>CDS version and<br>Driver Version | ClarityChrom®<br>CDS version and<br>Driver Version | Chromeleon™<br>CDS version<br>Driver version | ChromGate® CDS version and Driver Version | PurityChrom®     |
|--------------------------------------|--|--|--|---|------------------|
| Smartline Column Oven 4050           | n/a  | CC 2.4.4   | n/a  | CG 3.1.7 Driver 1466                      | n/a              |
| Memmert UN 55 Oven                   | n/a  | n/a  | n/a  | n/a                                       | Version 5.09.069 |
| Eluent heating device 6)             | n/a  | n/a  | n/a  | n/a                                       | Version 5.08.035 |

- 1) firmware version b1.17 or higher for JetStream required
- <sup>2)</sup> if connected, control panel must show "Choose Function" before the instrument will be opened
- <sup>3)</sup> device does not support static IP address, only LAN via DHCP from firmware version 1.11
- <sup>4)</sup> in PurityChrom version older than 5.09.134 only a thermostat or eluent heater is supported in one system
- 5) leak detector is not supported
- 6) separate software TempLogger for Standalone operation available

| Valves as fraction collector -> Fraction Collectors | OpenLAB® CDS<br>EZChrom Edition<br>CDS version and<br>Driver Version | ClarityChrom®<br>CDS version and<br>Driver Version | Chromeleon™<br>CDS version<br>Driver version            | ChromGate®<br>CDS version and<br>Driver Version | PurityChrom®      |
|---|--|--|---|---|-------------------|
| AZURA® Valve Unifier VU<br>4.1 1)                   | OLEE A.01.05 /<br>A.04.05<br>Driver 2138                             | CC7.4.1 Driver 5066                                | Cm 7.2 Regular Class<br>3 driver 1056 <sup>2), 3)</sup> | n/a 1)  | Version: 5.07.039 |
| AZURA® Valve V 2.1S <sup>1)</sup>                   | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                             | CC 3.0.7 Driver 3078 <sup>1)</sup>                 | Cm 7.2 Regular Class<br>3 driver 1056 <sup>2), 3)</sup> | n/a <sup>1)</sup>                               | Version: 5.07.039 |
| WellChrom electric valve<br>drives K-6, K-12, K-16  | n/a <sup>4)</sup>  | CC 2.4.4   | CM 6.8 Generic<br>driver <sup>3)</sup>                  | CG 2.8  | Version: 5.07.039 |

| Electrical Valve Drives (continued)  Valves as fraction collector -> Fraction | OpenLAB® CDS<br>EZChrom Edition<br>CDS version and<br>Driver Version | ClarityChrom® CDS version and Driver Version | Chromeleon™<br>CDS version<br>Driver version | ChromGate®<br>CDS version and<br>Driver Version | PurityChrom®      |
|---|--|--|--|---|-------------------|
| <u>Collectors</u>   |  |  |  |   |                   |
| Smartline valve drives S6,  | OLEE A.01.05 /   | CC 2.4.4                                     | CM 6.8 Generic                               | CG 3.1 1)                                       | Version: 5.07.039 |
| S12, S16 <sup>1)</sup>  | A.04.05  |  | driver 3)                                    |   |                   |
|   | Driver 1017  |  |  |   |                   |
| VICI Valco drives   | OLEE A.01.05 /   | CC 5.0.2 5)                                  | Implemented (not by                          | Generic Driver 3)                               | Version: 5.07.039 |
|   | A.04.05  |  | KNAUER)                                      |   |                   |
|   | Driver 1017  |  | ·  |   |                   |

<sup>-</sup> can be controlled as KNAUER switching valve if not listed in the module list of previous software versions (limited on RS-232 and only valve types available for V 2.1S)

| Other Supported Instruments                            | OpenLAB® CDS<br>EZChrom Edition<br>CDS version and<br>Driver Version | ClarityChrom® CDS version and Driver Version | Chromeleon™<br>CDS version<br>Driver version | ChromGate® CDS version and Driver Version | PurityChrom®      |
|--|--|--|--|---|-------------------|
| Advion expressION-S CMS MS <sup>1), 2)</sup>           | n/a  | ClarityChrom 5.0.3                           | n/a  | n/a                                       | n/a               |
| Microsaic 4000 MiD Single<br>Quad MS <sup>1), 2)</sup> | n/a  | ClarityChrom 7.4.1                           | -  | n/a                                       | Version: 5.09.069 |
| Airsensor 3)   | n/a  | n/a  | n/a  | n/a                                       | Version: 5.07.039 |
| Pressure control (pressure difference) 4)              | n/a  | n/a  | n/a  | n/a                                       | Version: 5.07.039 |

<sup>-</sup> generic driver for KNAUER switching valves will also work with AZURA® VU 4.1 (only for valve types available for V 2.1S, RS-232 only) and V 2.1S (RS-232 only), Cm 6.8 and 7.2

<sup>&</sup>lt;sup>3)</sup> - generic drivers do not support all of the device's features (e.g. RS-232 only, limited number of valve types)

<sup>-</sup> can be controlled as KNAUER switching valve

<sup>5) -</sup> driver was developed by DataApex

| Other Supported Instruments (continued)              | OpenLAB® CDS EZChrom Edition CDS version and Driver Version | ClarityChrom® CDS version and Driver Version | Chromeleon™<br>CDS version<br>Driver version | ChromGate® CDS version and Driver Version | PurityChrom®      |
|--|---|--|--|---|-------------------|
| Flowmeter GJC, Model<br>5025000 230V                 | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                    | n/a  | n/a  | supported                                 | n/a               |
| Flowmeter Bronkhorst Mini<br>Cori-Flow M12, M13, M14 | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                    | CC 8.1 driver 6018                           | n/a  | supported                                 | Version: 5.09.028 |

| Fraction Collectors (one fraction collector in a system, no combination of a fraction valve and a fraction collector possible)       | OpenLAB® CDS<br>EZChrom Edition<br>CDS version and<br>Driver version *) | ClarityChrom® CDS version and Driver version **) | Chromeleon™<br>CDS version<br>Driver version ***);<br>FRC license<br>required          | ChromGate® CDS version and Driver version ***); FRC license required | PurityChrom®      |
|--|---|--|--|--|-------------------|
| KNAUER Valve as Fraction<br>Collector (e.g. as<br>"MultiValve FC", all<br>KNAUER electric valve<br>drives and VU 4.1 <sup>2)</sup> ) | OLEE A.01.05 /<br>A.04.05<br>Driver 1017 <sup>1)</sup>                  | CCP 2.6 <sup>1)</sup>                            | Cm 7.2 Regular Class<br>3 driver 1056;<br>Refer Knauer FRC<br>manual for<br>Chromeleon | CG 3.1.6 Driver 423 <sup>1)</sup>                                    | Version: 5.07.039 |
| ASM 2.2L valve   | Coming soon   | CC 8.2.2 6036 <sup>14)</sup>                     | n/a  | n/a  | Version: 5.09.036 |
| ASM 2.1L valve   | OLEE A.01.05 /<br>A.04.05<br>Driver 1017 13)                            | CCP 3.0.7 Driver 3078 14)                        | n/a  | CG 3.3.2; Driver 2825  | Version: 5.07.039 |
| Smartline Fraction Collector 3050 4)   | n/a   | CCP 2.6  | n/a  | CG 3.3.1 Driver 1640   | n/a               |

driver was developed by DataApex
 device control requires MS license
 communication via game port adapter
 requires KNAUER IFU 2.1/ IFU 2.1 LAN

| Fraction Collectors<br>(continued)                | OpenLAB® CDS<br>EZChrom Edition<br>CDS version and<br>Driver version *) | ClarityChrom® CDS version and Driver version **), ***) 17)                           | Chromeleon™<br>CDS version<br>Driver version;<br>FRC license<br>required | ChromGate® CDS version and Driver version; FRC license required | PurityChrom®      |
|---|---|--|--|---|-------------------|
| Buechi Fraction Collector<br>C-684 <sup>5)</sup>  | n/a   | n/a  | n/a  | CG 3.1.6 Driver 423   | n/a               |
| Buechi Fraction Collector<br>C-660                | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                                | CCP 2.6  | n/a  | CG 3.1.6 Driver 423   | n/a               |
| ISCO Fraction Collector<br>Foxy Jr.               | n/a   | CCP 2.6  | Implemented (not by KNAUER)  | CG 3.1.6 Driver 423   | Version: 5.07.039 |
| ISCO Fraction Collector<br>Foxy R1 <sup>6)</sup>  | OLEE A.01.05 /<br>A.04.05<br>Driver 1017 7)                             | CCP 2.6.5 Driver<br>1094 (RS-232 only)<br>8)<br>CC 7.4.2 hotfix 5075<br>**) also LAN | Implemented (not by KNAUER)  | CG3.3.2 Driver 1751   | Version: 5.07.039 |
| ISCO Fraction Collector<br>Foxy R2 <sup>6)</sup>  | OLEE A.01.05 /<br>A.04.05<br>Driver 1017 7)                             | CCP 2.6.5 Driver<br>1094 (RS-232 only)<br>8)<br>CC 7.4.2 hotfix 5075<br>**) also LAN | CM 6.8 Generic<br>driver <sup>20)</sup>                                  | CG3.3.2 Driver 1751   | Version: 5.07.039 |
| ISCO Fraction Collector<br>Foxy 200 <sup>9)</sup> | n/a   | CCP 2.6 and<br>CCP 2.6.5 Driver<br>1100 9)   | Implemented  | CG 3.1.7 Driver 1466<br>-<br>CG3.3.2 Driver 1751                | n/a               |
| Virtual Fraction Collector                        | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                                | CCP 2.6  | n/a  | CG 3.1.6 Driver 423   | n/a               |
| Labocol Vario 2000<br>Fraction Collector          | OLEE A.01.05 /<br>A.04.05<br>Driver 1017                                | n/a  | n/a  | CG 3.1.6 Driver 423   | Version: 5.07.039 |

| Fraction Collectors<br>(continued)   | OpenLAB® CDS<br>EZChrom Edition<br>CDS version and<br>Driver version *)  | ClarityChrom® CDS version and Driver version **), ***) 17)   | Chromeleon™<br>CDS version<br>Driver version;<br>FRC license<br>required | ChromGate® CDS version and Driver version; FRC license required | PurityChrom®  |
|--|--|--|--|---|---|
| Labocol Vario 4000<br>Fraction Collector <sup>6)</sup>   | OLEE A.01.05 / A.04.05 Driver 1017 (RS-232 only) OLEE A.02.00 / A.04.06 Driver 2155 *) 18) also LAN; only fixed rack types | CC 7.4.2 Driver<br>5075 (hotfix) <sup>19)</sup><br>only fixed rack types   | -  | CG 3.3.1 Driver 1640<br>(RS-232 only)                           | Version: 5.07.039<br>(RS-232 and LAN<br>control, free rack<br>definition<br>supported, mixed<br>rack configuration<br>supported, no<br>USB control) |
| Labocol Vario 4000 Plus<br>Fraction Collector <sup>6)</sup>  | OLEE A.02.00 /<br>A.04.06 Driver 2155<br>*) <sup>18)</sup><br>only fixed rack types  | CC 7.4.2 Driver<br>5075 (hotfix) <sup>19)</sup><br>only fixed rack types   | -  | n/a   | Version: 5.07.039<br>(RS-232 and LAN<br>control, free rack<br>definition<br>supported, mixed<br>rack configuration<br>supported, no<br>USB control) |
| VICI Valco valve as VICI<br>MultiValve Fraction<br>Collector   | OLEE A.01.05 /<br>A.04.05<br>Driver 1017 <sup>1)</sup>   | n/a from KNAUER,<br>in ASM 2.1L as FC <sup>3),</sup><br>10)<br>valve as FC is<br>supported by<br>DataApex <sup>11)</sup> | n/a  | CG 3.3.1 Driver 1640  | Version: 5.07.039   |
| Injection Module<br>(combined module with<br>pump and injection valve<br>for injection of large<br>injection volumes, allows<br>for stacked injection) 12) | OLEE A.01.05 /<br>A.04.05<br>Driver 1017   | n/a  | n/a  | CG 3.3.1 Driver 1640  | Version: 5.07.039   |

- \*) Due to changes in OpenLAB EE by Agilent, the KNAUER FRC AddOn needed to be adjusted and is not compatible to OpenLAB EE A.01.05 / A.04.05 from driver package version 2146. All later added functionality and drivers are not available for OpenLAB EE A.01.05 / A.04.05.
- \*\*) Up to version 6.1, all KNAUER preparative functionality and fraction collector drivers are only available in ClarityChrom *Prep*. From version 7.2., the ClarityChrom *Prep* as a separate software is discontinued. The full functionality now is available in ClarityChrom with the special Knauer FRC control. If an upgrade of ClarityChrom *Prep* is required, a ClarityChrom with Knauer FRC control will be delivered (new User Code).
- \*\*\*) The listed fraction collectors are supported in the KNAUER FRC module for ClarityChrom (from version 7.4.2) or in former ClarityChrom® *Prep*. DataApex supports fraction collection with an own implementation without a separate license; it is available in ClarityChrom® as well as ClarityChrom® *Prep*.
  - 1) up to 32 valves in a cascade
  - <sup>2)</sup> VU 4.1 as MultiValve FC fully supported from CC 7.4.1.5066, in previous versions as V 2.1S RS-232 only
  - <sup>3)</sup> only one valve can be used as fraction collector, no cascading is supported
  - 4) switch between RS-232 and LAN with a switch on the devices rear panel
  - <sup>5)</sup> requires special relay module for serial communication
  - 6) touch panel is not blocked in remote, do not touch it while the device is under software control
  - LAN control from firmware version 1.01.03, LAN control support in firmware versions 1.01.09 and 1.01.11 faulty
  - 8) from CCP 2.6.5 1094 only RS-232 support, from 7.4.2 hotfix 5075 also LAN control
  - 9) support is discontinued due to several incompatible firmware versions; not all devices may work with the driver
  - <sup>10)</sup> not all available valve types can be built-in into ASM 2.1L
  - <sup>11)</sup> driver from DataApex available; does not support KNAUER fractionation features
  - several pumps were added to the module when released; however, a pump is only supported in the injection module if the mentioned pump is also available as a separate pump; in ASM 2.1L pump and valve must be controlled separately, refer software manual for detailed information
  - only one valve as fraction collector, no cascading if controlled as ASM; FRC valve cascading only is possible if a valve in the ASM is configured as KNAUER MultiValve FC and is controlled as separate device via IP port; more information about control as separate device via IP port in software manual
  - only one valve as fraction collector in ASM, no cascading
  - only one valve as fraction collector, no cascading if controlled as ASM; FRC valve cascading only is possible if a valve in the ASM is configured as KNAUER MultiValve FC and is controlled as separate device via IP port
  - <sup>16)</sup> only one valve as fraction collector, no cascading
  - <sup>17)</sup> DataApex supports fraction collection with an own implementation; it is available in ClarityChrom® as well as ClarityChrom® *Prep*. KNAUER FRC option is available in ClarityChrom® *Prep* and, from version 7.4.1, also in ClarityChrom® with Knauer FRC license

- from driver version 2155, additionally model 4000 Plus, LAN control and trays with 24 and 39 vials are supported, now trays with 20, 24, 39, 80 and 125 vials are available, free tray definition is not supported; all racks must have same type (no mixed configuration supported); no USB control support; former drivers only support model 4000, RS-232 control and trays with 20, 80 and 125 vials
- only trays with 20, 24, 39, 80 and 125 vials are supported; free tray definition is not supported; all racks must have same type (no mixed rack configuration supported); no USB control support
- <sup>20)</sup> generic driver should also work with Cm 7, generic drivers do not support all of the device's features (e.g. RS-232 only)

#### **Mobile Control - KNAUER**

Mobile Control requires minimum device firmware versions, which may differ between the Mobile Control versions. Refer to the list of supported devices and their required firmware version below. A Mobile Control update may also require firmware updates of the controlled devices. Mobile Control download comes with a Firmware Wizard, which allows for firmware updates. For **Monitor Mode, supported in OpenLAB® EE and ClarityChrom® only**, every Mobile Control version requires a minimum KNAUER driver version for the respective CDS.

**Mobile Control, Version 5.5.36** also with data acquisition support, supports AZURA® devices:

- Pump P 2.1S/P 4.1S Firmware V01.38
- Pump P 6.1L Firmware V01.06
- Pump P 2.1L Firmware V01.09
- Detector UVD 2.1S Firmware V01.11
- Detector UVD 2.1L Firmware V02.06
- Detector MWD 2.1L Firmware V01.10

- Detector DAD 2.1L\* Firmware V01.10
- Detector DAD 6.1L\* Firmware V01.23
- Detector RID 2.1L Firmware V01.24
- Valve V 2.1S Firmware V05.01
- VU 4.1 Firmware V05.68
- Assistant ASM 2.1L Firmware V01.17

- Assistant ASM 2.2L Firmware V01.10
- Column thermostat CT 2.1 Firmware V01.06
- Autosampler AS 3950 Firmware V01.17
- Autosampler AS 6.1L Firmware V01.17
- IFU 2.1 LAN Firmware V1.05

Mobile Control version 5.5.36 supports Monitor Mode in OpenLAB® EE from driver version 2155 or in ClarityChrom® from version 8.1. Monitor Mode is not supported in any other CDS, e.g., PurityChrom® and Chromeleon™. Either CDS or Mobile Control can access the devices.

<sup>\* -</sup> no 3D data acquisition, 4 channels only

#### **EuroChrom for Windows 3.05 (KNAUER)**

Note that EuroChrom was discontinued in 2008. It will not run on Windows systems newer than Windows XP. Fraction collectors are supported exclusively in EuroChrom Prep software.

#### <u>Pumps</u>

Smartline Pump 1000, Preparative Pump 1800, WellChrom K-1800, WellChrom K-1001, K-1000, K-501, K-120 (Smartline 100 as K-120 only via RS-232 without pressure sensor),

Feed Delivery (ECW Prep only): pump K-501, K-120

#### **Detectors**

Smartline UV Detector 2500, Smartline RI Detectors 2300 / 2400, Smartline DAD 2800 (PCI/LAN), WellChrom UV Detectors K-2000/2001, K-2500/2501, K-2600, K-200, WellChrom RI Detectors K-2301/2401, WellChrom DAD K-2700 (PCI/LAN), WellChrom DAD K-2800 (PCI/LAN), Virtual Detector

(WellChrom detectors 2xx0 allow only data acquisition, no control)

## <u>Autosamplers</u>

AS 3800 / Spark Basic Marathon, AS 3900 / Spark Midas, Spark Triathlon

# **Analog/Digital Converters**

KNAUER IF2 (also IFU 2.1 with built-in USBtoCOM converter), Smartline Manager 5000, KNAUER HPLC Box (Model 96)

# Thermostats/Column Ovens

JetStream (FW b1.17 or later required)

#### <u>Valves</u>

KNAUER WellChrom K-6 / K-12 / K-16 [Smartline S 6 / S 12 / S 16 and AZURA® V2.1S and VU 4.1 (only 2, 6, 12, 16 port valves) are compatible via RS-232],

Solvent Recycling: K-6 / K16

## Fraction Collectors (ECW Prep 3.05 only)

Labocol Vario 2000, ISCO Foxy Jr, KNAUER WellChrom K-6 / K-12 / K-16 (also cascade up to 8 valves), Buechi 684 via short circuits, User Defined FC (via short circuits)

#### **Other Software**

**OpenLAB CDS** (Agilent) - currently no drivers from KNAUER available

OpenLAB CDS ChemStation Edition (Agilent) - no drivers from KNAUER are available

**LabVIEW** (National Instruments) - <u>support is discontinued</u>, may not work with LabVIEW versions higher than 2011

Driver for AZURA® Pumps P 2.1S / P 4.1S, Smartline 100, BlueShadow Pumps 10P / 20P

# **HyStar™** (Bruker)

Compass HyStar<sup>™</sup> 3.2 Plug-in: Drivers for PLATINblue system (Pump P-1, DAD PDA-1, UV Detector MW-1, Column Thermostat T-1, Autosampler AS-1, Manager M-1 (LPG valves only, interface module is not supported)

Compass HyStar<sup>™</sup> 4 Plug-in: Driver for AZURA® DAD 6.1L

## © KNAUER All rights reserved.

Version 1.18 Last update: 2021/03/23

KNAUER Wissenschaftliche Geräte GmbH Hegauer Weg 38 14163 Berlin, Germany

Tel.: +49-(0)30-809727-0 Fax: +49-(0)30-8015010 E-mail: info@knauer.net Internet: www.knauer.net

#### **Xcalibur™** (Thermo)

Version 2.0.7 and 2.1: Driver for PLATINblue Pump P-1, UV Detector MW-1, DAD PDA-1, Autosampler AS-1, Column Thermostat T-1 and T-1 Basic and Smartline Pump S 1050, UV Detector S 2550, Autosampler S 3950 and Column Thermostat JetStream

#### **TRADEMARKS**

The company, product and service names used in this brochure are for identification purposes only. All trademarks and registered trademarks are the property of their respective owners.

-----

This document is subject to technical changes. Please visit <a href="www.knauer.net">www.knauer.net</a> for the latest version, ordering details and overview of article numbers.